



Biochemical Diagnostics, Inc.

A Kova International Company

180 Heartland Blvd, Edgewood, NY 11717 • Phone (800) 223-4835

Fax (631) 595-9204 • www.biochemicaldiagnostics.com

CERTIFICATE OF ANALYSIS

**PRODUCT NAME: DETECTABUSE® STAT-SKREEN® LIQUID CONTROL URINE
with MDMA, OXYCODONE & BUPRENORPHINE**

2X Cutoff

CATALOG # 19001066 – 5 mL / 19002066 – 20 mL

Verified Value (ng/mL)

CONSTITUENTS	LOT #: BK00027 EXP: 2020-04-30	
	Target	Verified
SAMHSA Mandated		
DELTA-9-THC-COOH	100	96
BENZOYLECGONINE	600	580
PHENCYCLIDINE (PCP)	50	53
MORPHINE	600	603
d-METHAMPHETAMINE	2000	2003
d-AMPHETAMINE	2000	1985
Non-Mandated		
SECOBARBITAL	600	623
OXAZEPAM	600	597
METHADONE	600	621
METHAQUALONE	600	597
d-PROPOXYPHENE	600	620
NORTRIPTYLINE	2000	2040
MDMA	1000	982
OXYCODONE	200	196
BUPRENORPHINE	20	20

Biochemical Diagnostics, Inc., is an FDA licensed manufacturing facility and as such, we manufacture all in vitro diagnostic products under the guidelines of current GMP and other applicable governing regulations. In addition, all of our liquid control urine products are manufactured in compliance with the directive 98/79/EC of the European Parliament and of the Council and carry the CE mark.

Biochemical Diagnostics Liquid Control Urine is manufactured from negative, human based urine spiked with the specified constituents. All drug standards used in manufacturing are at least 98% minimum purity. Specific gravity, pH and creatinine fall within the normal limits of human urine. Contains 0.05% sodium azide.


Controls are assayed by SAMHSA and/or CAP certified laboratories verified by GC/MS and/or LC/MS which, based on our experience, provide consistent, reliable results with predictable statistical bias within their acceptance criteria.

Because SAMHSA certified laboratories are permitted to accept values that fall within $\pm 20\%$ of assigned control values for SAMHSA mandated drugs, and an often looser criteria for non-mandated drugs, your assay values, or those obtained from other outside laboratories may be biased at either extreme of their acceptance criteria and may not match the BCD values.

Please contact us for guidance regarding these issues. For storage, stability or additional information refer to package insert.

This product passes our in-house quality control release specifications.

QA Approval:


11/20/17